

Autoclave Cycle Verification Testing Using Biological Indicator Ampoules



Human Resources, Occupational Health & Safety,
University of Western Ontario, October 2003



Tests using biological indicator ampoules to prove that autoclaving has rendered the laboratory waste non-hazardous must be conducted at least every 6 days or before any waste is placed into the municipal waste stream as specified in Guideline C - (Non - Incinerated Technologies for Treatment of Biomedical Waste, Ministry of the Environment, Ontario, December 2001).

INSTRUCTIONS:

