



LEVEL 3 CONTAINMENT INSPECTION REPORT

DATE: MARCH 23, 2009
INSPECTED BY: JENNIFER STANLEY

The Level 3 containment unit was inspected on March 23, 2009. The facility was clean and neat and in good order. The following items were observed:

1. The Rugar project still needs to have a Biosafety sign on it (ACVS Room 6009). I will send a Level 2 sign for posting.
2. The certification reports for the biological safety cabinets have been posted on the bulletin board in the hallway of the facility.
3. The WSIB has updated the First Aid Requirements, Regulation 1101. The updated version of this regulation has been placed in both first aid kits (found in the hallway and in Room 6005A).
4. The florescent lights and green indicator light for the fan have been fixed/replaced.
5. The flammable substances have been removed from the laboratory and are stored in the flammable cabinet in Room 6011. Thank you for complying with this requirement.